



Range 0.01 - 4 mmol/L


Linearity Linearity at high range: 4 mmol/L




Linearity at low range:



Correlation



Sensitivity 0.005 mmol/L



Interference Ascorbic acid and hemolysis do not have significant effects on the assay. Citrate, oxalate, EDTA and sodium fluoride do not have significant influences on the assay when they are used in their usual amounts.

CE Applications	AU400	Hitachi 902	Hitachi 912
	AU600	Hitachi 904	Hitachi 7600
	AU640	Hitachi 911	Modular

Ordering	Code No.	Product	Content
	434-91795	NEFA-HR(2) R1 Set	R1: 4 x 50 mL R1a: 4 x for 50 mL
	436-91995	NEFA-HR(2) R2 Set	R2: 4 x 25 mL R2a: 4 x for 25 mL
	270-77000	NEFA Standard	CAL: 2 x 10 mL